

Publishable Summary for 22NRM05 MeLiDos

Metrology for wearable light loggers and optical radiation dosimeters

Overview

New standardised metrics for measuring the physiological, non-visual effects of light have been recently introduced in CIE S 026 and wearable light loggers (also known as light dosimeters) have emerged as a way to measure a person's light dose. However, the optical characterisation of these dosimeters remains challenging because of the compromises needed to make them wearable and affordable. This project aims to provide the scientific and industrial communities with the necessary tools (e.g. best practice guides) for the characterisation and use of wearable light loggers.

Need

Daylight affects human physiological activity and regulates many aspects of society (e.g. waking hours, alertness). On the other hand, the invention of electric lighting has led people to spending more time indoors during the day, including working remotely from home, and staying up later in the evening. However, the intensity of artificial light is only a fraction of that of daylight (a half for a cloudy sky, and less than 1 % of a sunny afternoon), while their spectral compositions differ in terms of blue light proportion. This point is of particular importance because of melanopsin, a photo-pigment present in the eye and involved in physiological responses to light but not in direct vision, called Non Image-Forming (NIF) effects. Implications for health and well-being vary from mood and wakefulness to sleep deprivation, including even more severe health problems, such as heart diseases and cancer. While the study of lighting sources can be performed in laboratories under controlled conditions, the light humans receive under actual environmental conditions is very different due to e.g. window placement, reflectance of materials in the surrounding, habits, etc.

Wearable light loggers are currently being used for the measurement of light exposure on humans, as they are wearable and provide light exposure data during the course of a whole day. These light dosimetry time series are then correlated to other factors, such as melatonin levels or heart rate, to investigate the physiological effects. Although they are compact and lightweight, they often have a black-box design where the sensor technology (e.g. filter-based or spectrally resolved), the data processing (e.g. sampling rate, averaging or other data processing, etc.), or both, are unknown, which makes evaluating their performances even harder. It can then lead to inaccuracies, inappropriate metrics or conclusions due to misinterpretation or incomplete data. There are currently no guidelines or standards for the characterisation of light loggers, and the collected data are thus often considered unreliable due to inaccurate device specifications and/or data processing or interpretation (e.g. a user not familiar with the field misinterprets a specification or metric).

In TN002:2014, CIE has identified the need for the photobiological and photometric quantities to be harmonised. With the rapidly evolving field of lighting (future revision of CIE S 025, publication of CIE S 026, introduction of daylighting by EN 17037 and revision of workplace lighting in EN12464-1) and the spread of wearable devices of all kinds, it is important to keep metrology recommendations and guidelines up to date to ensure reliable studies well support health recommendations and policies.

Objectives

The overall aim of the project is to introduce metrology for light dosimeters, and in particular for light loggers designed to respond to the needs raised by CIE S 026 for measuring melanopsin-related metrics. More

Report Status:
PU – Public, fully open

Funded by the European Union. Views and opinions expressed are however those of the author(s) only and do not necessarily reflect those of the European Union or EURAMET. Neither the European Union nor the granting authority can be held responsible for them.

European Partnership



Co-funded by
the European Union

Publishable Summary

The project has received funding from the European Partnership on Metrology, co-financed from the European Union's Horizon Europe Research and Innovation Programme and by the Participating States.

**METROLOGY
PARTNERSHIP**



Issued: June 2023

specifically, this project will deliver the metrics and procedures, as well as guidelines based on use cases, in order to lay the groundwork addressing the normative standards' needs.

The specific objectives are:

1. To develop and validate new characterisation methods, and define a validated set of quality indices for wearable optical radiation dosimeters, with a focus on ranges relevant to daily life (i.e. UV and visible). Their validation will be performed against lab-based instrumentation for parameters (e.g. wavelength sensitivity, linearity and angular response) in an intercomparison exercise. The quality indices will be based on those defined by ISO and CIE for illuminance meters and luminance meters.
2. To develop data analysis methods to verify and increase the quality of data, as well as to establish the correlation between different datasets to improve reproducibility and make comparisons from different devices more reliable.
3. To analyse use cases for wearable light loggers based on end user needs, and to then produce recommendations based on those needs on i) (meta-)data model and data formats, as well as summary metrics and on ii) the practical use of light loggers, including device placement, compliance and quality control.
4. To evaluate the performance of new portable devices able to measure spatially resolved photobiological quantities and to compare them with spatially averaging devices, to determine relevancy depending on use case. Additionally, to investigate the shortcomings of current devices regarding their use and to investigate the benefits and drawbacks of spatially-resolved devices.
5. To collaborate with CIE Division 2 and Division 6, and the users of the standards they develop to ensure that the outputs of the project are aligned with their needs, including the provision of recommendations in a form that can be incorporated into future standards at the earliest opportunity. Additionally, to promote the results to other relevant parties such as lighting experts, chronobiologists and sleep specialists through various communication channels available to the participants.

Progress beyond the state of the art and results

Development of characterisation methods and quality indices for light loggers and optical/UV radiation dosimeters (Objective 1):

Though it has been known for over two decades that light has a profound effect on human health and wellbeing, only recently has CIE S 026 defined spectral sensitivity functions, quantities, and metrics to quantify the light triggering these effects. The spectral measurement and characterisation of light sources in testing laboratories are usually well understood and the measurement uncertainties are at an acceptable level for current uses. In recent years, various wearable light loggers and UV radiation dosimeters have been developed to enable the measurement of spectrally weighted irradiances and irradiance doses in field experiments, implementing the standard CIE S 026. Since these instruments need to be portable and light weight, their design differs from those typically used in laboratory environments. Moreover, since these instruments are to be used in field experiments and deployed in numbers, the evaluation of their performance needs to be cheap and time-efficient. Traditional methods of providing traceability for example by measuring the spectral responsivity with the narrow band source are thus not suitable. So far, no standardised methods exist to characterise wearable light loggers and UV radiation dosimeters.

This project will build a metrological framework for the characterisation of wearable light loggers, giving methods for their evaluation, validation, and traceability. The developed methods will be validated through an intercomparison, contributing to measurement capabilities at NMIs as well as in industry. Performance indicators, consisting of a set of quality indices, will be developed in order to make it easier for users to assess the performance of a device in terms of e.g. linearity. The project outcomes will be summarised in a publicly available good practice guide (GPG), which will support CIE in the development of standards for the characterisation of wearable light loggers. The overall goal is to provide guidance and methods to stakeholders on how to reliably characterise wearable light loggers.

Development of data analysis methods for data quality verification and data reproducibility (Objective 2):

The data currently collected by wearable light loggers, are usually affected by multiple factors, for example who is wearing them (i.e. if the device is worn correctly, or at all times, etc.), the indoor/outdoor environment,

etc. It is therefore necessary to include processes and data analysis methods to validate the data. This project will provide the necessary guidelines for preparing and carrying out field studies with wearable light loggers in a way that promotes data quality and reliability, with a user-oriented mind-set. This will include data and metadata formats and data validation tools and recommendations to facilitate the reproducibility, data transfer and comparison between data sets. A research-oriented software ecosystem will be included, comprised of various modules for data format conversion, data analysis, validation, and visualisation, as well as a cross-comparison between different devices and studies with harmonised data format.

To analyse use cases for wearable light loggers based on end user needs, and to then produce recommendations based on those needs on i) (meta-)data model and data formats, as well as summary metrics and on ii) the practical use of light logger (Objective 3):

There are multiple applications for wearable light loggers and different devices designed to answer specific applications. For example, a study on the effect of light sensitivity on ASD (Autistic Syndrom Disorder) patients will have different requirements to a study of wall paint colour on wakefulness in office workers, both in terms of device technical specifications and in terms of practical use (e.g. when, how the device is worn, data relevancy, etc.). There are currently no guidelines or recommendations related to the proper selection and use of wearable light loggers. This project will define a usage landscape based on the applications' needs and provide recommendations on the quality indices and other characteristics that a user should pay attention to while choosing and using wearable light loggers.

Portable devices performance evaluation for the measurement of spatially resolved photobiological quantities. Comparison with spatially averaging devices and investigation of benefits/drawbacks of spatially-resolved devices (Objective 4):

The fast development of wearable devices and the spread of their use for aggregating data with constellations of devices or for behavioural studies, poses a major challenge for the reliability and traceability of the data generated by these devices. New sensor technologies for wearable light loggers will be investigated in this project in order to ensure the delivered methods and recommendations will remain transferrable to new technologies. In particular, the use of spatially resolved devices will be investigated since the eye is also a spatially resolved sensor and the study of non-visual effects of light should not stop at spatially averaged light levels.

Outcomes and impact

Outcomes for industrial and other user communities

The project will provide the industrial and scientific communities in the field of wearable light loggers and UV radiation dosimeters with the methods, recommendations, and guidelines as necessary for the characterisation of these devices. The quality indices will not only benefit users as a way to assess the performance of a device but also industry that wants to design improved devices or similar devices that look at other wavelengths.

Improving light dosimetry will also enhance the quality of studies regarding Non-Image Forming (NIF) effects of light and benefit lighting experts, interior architects and similar professions by providing them with useful tools to evaluate their work in situ. As both regulations and public awareness are progressing towards more health considerations in buildings, it will provide opportunities for the lighting industry to improve their products to be more in line with trustful, state of the art research. Light loggers could also be used by these experts, empowering them with new tools to easily monitor the lighting changes throughout a day or with changes in furniture, paint, and other interior arrangements.

Outcomes for the metrology and scientific communities

The project will deliver recommendations and guidelines for the characterisation and use of wearable light loggers. This will improve the quality and reliability of the data generated by clinical studies, as well as the robustness of their results and conclusions. Researchers and clinical staff will be able to choose devices which best fit their application as well as get more reliable light dosimetry data and therefore, more reliable conclusions in their studies. As the project includes efforts to harmonise data format and analysis methods, as well as a research-oriented software ecosystem (i.e. data formatting and conversion, analysis, validation and visualisation packages), it will also enhance reproducibility of research or clinical studies by providing a set of open-access tools for data analysis and representation, as well as for cross comparisons between studies.

While work on solar UV radiation dosimeters will not be as extensive in this project, it will still benefit scientists by providing them with more reliable tools to measure UV doses in the field and assess its impacts on human health.

The project will also improve the performance of the wearable light loggers that will be designed for and used by the general public, providing more exhaustive and accurate information about the light dose they receive. Data provided by the general public (in this case light dosimetry data) can be used for Big Data studies. Making this light dosimetry Big Data reliable is again key for the researchers that will use it to make statistical analyses of practices in the general public.

The metrology benefits are more straightforward, as the project will lay the groundwork for normalisation bodies regarding wearable light loggers, with the methods not only limited to wearable light loggers for NIF effects of light, but also including UV radiation dosimeters.

Outcomes for relevant standards

The results of this project will be used to complement CIE S 026, which defines the metrics for NIF effects of light, with recommendations on how these should be properly measured, such as device requirements, characterisation and practical use. The project will also contribute to the implementation of standards in daylighting EN 17037 and workplace lighting in EN12464-1 by improving the reliability of the tools available for building experts to assess their compliance to those standards. As currently there is no European standard on solar UV exposure for outdoor workers, this project should contribute to the development of new standards on solar UV exposure, in particular by providing insights for potential Personal Prevention Equipment (PPE) against solar UV radiation.

Longer-term economic, social and environmental impacts

Currently, very little is known about how light exposure affects human health and wellbeing. The introduction of metrology for wearable light loggers will ensure scientists and clinical staff have reliable data to better understand the role of lighting in human health and wellbeing. Chronobiologists and sleep specialists will benefit from having reliable tools to study and understand the effect of light exposure on human physiology. The confidence in scientific results mentioned above will hasten the emergence of efficient guidelines and standards for the lighting field as well as health experts. This in turn will improve the livelihood, wellbeing and productivity of the European and other communities for the long term. The same health benefits are expected for the extended work to UV radiation dosimeters. The emerging increase in UV exposure as a result of global warming is expected to increase the number of skin cancer cases, as warmer and sunnier days means more time spent exposed to UV radiation. In view of this, characterisation methods and quality indices for the performance of solar UV dosimeters with the aim of optimising protective measures and effective strategies for skin cancer prevention are of great importance.

Moreover, connected devices for monitoring various parameters are already available to the general public (Internet of Things or IoT) but it is not always known what the data means exactly, both among the general public and sometimes, the manufacturer. Recommendations and future standards for the characterisation of light dosimeters, as well as recommendations for the formatting and use of the data, will improve the understanding of the health implications of light exposure by manufacturers, and in turn by the end user.

List of publications

-

This list is also available here: <https://www.euramet.org/repository/research-publications-repository-link/>

Project start date and duration:		01 June 2023, 36 months	
Coordinator: Fabien Eloi, LNE		Tel: +33130691340	E-mail: fabien.eloi@lne.fr
Project website address: https://www.melidos.eu/			
Chief Stakeholder Organisation: Commission Internationale de l'Eclairage		Chief Stakeholder Contact: Kathryn Nield, CIE General Secretary	
Internal Beneficiaries:		External Beneficiaries:	Unfunded Beneficiaries:
1. LNE, France		7. FUSP-CEU, Spain	13. BAuA, Germany
2. Aalto, Finland		8. Iberoptics, Spain	
3. CMI, Czechia		9. INEGI, Portugal	
4. RISE, Sweden		10. THUAS, Netherlands	
5. TUBITAK, Türkiye		11. TUM, Germany	
6. VSL, Netherlands		12. UCM, Spain	
Associated Partners:			
14. HSLU, Switzerland			
15. METAS, Switzerland			
16. SFI Davos, Switzerland			